

PRODUCT DATA SHEET

Standing: 2023-02-17

LAUDA Variocool VC 7000

Process thermostat 400 V; 3/N/PE; 50 Hz

Part Number: L000951

Features

- Process thermostat suitable for use with non-flammable heat transfer liquids
- Coloured TFT display for simultaneous indication of actual & set values and graphic illustration of the temperature profile
- Clear text menu navigation, six selectable languages DE, EN, FR, ES, IT, RU
- Easy input via cursor and soft keys
- Fully electronic continuous controller with PID action
- Electronic level indication and low level alarm
- Powerful pressure pump
- USB interface as standard
- Remote fault indication through floating contact
- Upgradeable with an interface module (analogue module, contact module, RS 232/485 module, Profibus, Ethernet-USB module)
- Integrated programmer with max. 150 segments, splittable in 5 programmes
- Adjustable bypass for pressure limiting
- Filler opening on top, drain tap on the backside
- SmartCool system for energy-saving digital cooling management including compressor on-off control
- Operates with non flammable liquids (water, water/glycol)
- Condenser cooling Air



Reserve technical changes



Working temperature min.

-20 °C



Working temperature max.

80 °C

LAUDA DR. R. WOBSEY GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0 • F + 49 (0) 9343 503-222
info@lauda.de • www.lauda.de
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Ralf Hermann, Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser

PRODUCT DATA SHEET

Standing: 2023-02-17

LAUDA Variocool VC 7000

Process thermostat 400 V; 3/N/PE; 50 Hz

Part Number: L000951

Technical Features (according to DIN 12876)

| | |
|---|---|
| Working temperature range | -20 ... 80 °C |
| Ambient temperature range | 5 ... 40 °C |
| Temperature stability | 0.1 ± K |
| Heater power max. | 4.5 kW |
| Power consumption max. | 8.8 kW |
| Current max. | 15 A |
| Pump Pressure max. | 5,0 bar |
| Pump flow rate max. (pressure) | 60 L/min |
| In / Outlet connection thread (outside) | G 1 1/4" |
| Pressure adjustment | bypass |
| Filling volume max. | 64 L |
| Overall dimensions (WxDxH) | 650 x 670 x 1250 mm |
| Weight | 124 kg |
| Refrigerant stage 1 | R-452A (GWP 2140); 2.000 kg; 4.3 t CO2-eq |
| Power supply | 400 V; 3/N/PE; 50 Hz |
| Power plug | Power cord with plug (IEC 60309, 5-pol, CEE, red, 16 A) |

Reserve technical changes

| Temperature | Heat transfer liquid | Cooling Capacity 50 Hz |
|-------------|----------------------|------------------------|
| 20 °C | Ethanol | 6.65 kW |
| 10 °C | Ethanol | 4.95 kW |
| 0 °C | Ethanol | 3.35 kW |
| -10 °C | Ethanol | 2.05 kW |
| -20 °C | Ethanol | 0.95 kW |

LAUDA DR. R. WOBSEER GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0 • F + 49 (0) 9343 503-222
info@lauda.de • www.lauda.de
WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEER Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Ralf Hermann, Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser

PRODUCT DATA SHEET

Standing: 2023-02-17

LAUDA Variocool VC 7000

Process thermostat 400 V; 3/N/PE; 50 Hz

Part Number: L000951

Standard accessories

- 2 nipples 1" with screw cap G1/4 for pump connectors

LAUDA DR. R. WOBSEY GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0 • F + 49 (0) 9343 503-222
info@lauda.de • www.lauda.de
WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Ralf Hermann, Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser