

PRODUCT DATA SHEET

Published: 2026-01-08

Initial validation certificate for Versafreeze
one measuring point, at max. control temperature
Part Number: D000095

Features

- Initial validation certificate for the temperature adjustment of the device at maximum control temperature.
- Carried out at the factory as part of the regular routine testing (1-point measurement next to the standard sensor) with a calibrated measuring section.
- Creation of a confirmation including a copy of the calibration certificate for the measuring equipment.

LAUDA DR. R. WOBSEY GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0
info@lauda.de • www.lauda.de
WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser