

## PRODUCT DATA SHEET

Published: 2026-04-09

### LAUDA Microcool MC 350

Circulation chiller 220 V; 60 Hz / 230 V; 50 Hz

Part Number: L004454

#### Features

- Circulation chiller
- 1-line LED display for display of actual or set temperature
- User-friendly menu guidance with simplest 3-key operation
- Fully electronic continuous controller
- Temperature protection for the pump windings, lighted level indication and low level alarm
- Magnetically coupled block pump
- RS 232 interface
- Auto-Start-Timer and Auto-Shut-Down function
- Filler opening on top, outlet connection on the backside
- Cooling capacity adjustment by solenoid valve control including compressor on-off control
- Utilizes natural refrigerants
- Condenser cooling Air
- Operates with non flammable liquids (water,water/glycol)



Reserve technical changes



Working temperature min.  
-10 °C



Working temperature max.  
40 °C

LAUDA DR. R. WOBSEY GMBH & CO. KG  
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0  
info@lauda.de • www.lauda.de  
WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim • HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser

## PRODUCT DATA SHEET

Published: 2026-04-09

LAUDA Microcool MC 350


Circulation chiller 220 V; 60 Hz / 230 V; 50 Hz

Part Number: L004454

### Technical Features (according to DIN 12876)

|   |  |
|---|--|
| Working temperature range               | -10 ... 40 °C                                      |
| Ambient temperature range               | 5 ... 40 °C  |
| Temperature stability                   | 0.5 ± K  |
| Pump Pressure max.                      | 0,4 bar  |
| Pump flow rate max. (pressure)          | 16 L/min   |
| In / Outlet connection thread (outside) | Ø10 mm   |
| Filling volume min.                     | 4 L  |
| Filling volume max.                     | 7 L  |
| Overall dimensions (WxDxH)              | 240 x 400 x 500 mm                                 |
| Weight                                  | 30 kg  |
| Noise level                             | 57 dB(A)   |
| Net Weight                              | 30.4 kg  |
| Power consumption max.                  | 0.34 kW  |
| Current max.                            | 1.7 A  |
| Refrigerant stage 1                     | R-290 (GWP 3); 0.041 kg; 0.0 t CO <sub>2</sub> -eq |
| Power supply                            | 220 V; 60 Hz / 230 V; 50 Hz                        |
| Power plug                              | Power cord with plug (GB2099, 15934)               |

Reserve technical changes



Power cord with plug (GB2099, 15934)

LAUDA DR. R. WOBSEY GMBH & CO. KG  
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0  
info@lauda.de • www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim • HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser

## PRODUCT DATA SHEET

Published: 2026-04-09

LAUDA Microcool MC 350

Circulation chiller 220 V; 60 Hz / 230 V; 50 Hz

Part Number: L004454

| Temperature | Heat transfer liquid | Cooling Capacity 50 Hz | Cooling Capacity 60 Hz |
|-------------|----------------------|------------------------|------------------------|
| 40 °C       | Waterglycol          | 0.35 kW                | 0.35 kW                |
| 20 °C       | Ethanol              | 0.35 kW                | 0.35 kW                |
| 10 °C       | Ethanol              | 0.27 kW                | 0.27 kW                |
| 0 °C        | Ethanol              | 0.2 kW                 | 0.2 kW                 |
| -10 °C      | Ethanol              | 0.12 kW                | 0.12 kW                |

Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG  
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0  
info@lauda.de • www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim • HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser


## PRODUCT DATA SHEET

Published: 2026-04-09

LAUDA Microcool MC 350

Circulation chiller 220 V; 60 Hz / 230 V; 50 Hz

Part Number: L004454



Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG  
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0  
info@lauda.de • www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim • HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser