

## PRODUCT DATA SHEET

Published: 2026-04-28

LAUDA Universa U 830 E

Cooling thermostat 100 V; 50/60 Hz

Part Number: L003978

### Features

- Bath circulation thermostat for standard temperature control tasks up to 100 °C – reliable and intuitive to use.
- Future-proof thanks to modularity: LAUDA Universa with interchangeable control head and heating/cooling base allows flexible adaptation to current and future requirements.
- Bright VA LC display with clear two-line display and intuitive 3-button operation.
- Reliable circulation pump with flow reduction for adjusted volume flow and constant pressure.
- Energy-efficient refrigeration system: capillary tube technology with speed-controlled inverter compressor and speed-controlled fan - for reduced energy costs and a long service life.
- Environmentally friendly refrigerants: Use of natural refrigerants for sustainable and environmentally conscious cooling.
- Integrated timer with countdown function for time-controlled temperature control processes.
- Intelligent remote control and monitoring: Integrated web server allows flexible remote control in the company network via the LAUDA Command app or browser-based, secured by PKI and 2-factor authentication. The connection to LAUDA.LIVE enables global, cloud-based data analysis and remote maintenance.
- Modern connectivity thanks to standard integrated Ethernet and USB interfaces for reliable data communication.
- Wireless communication: Integrated WLAN enables effortless connection with PC, tablet and smartphone.
- Proven safety features: low-level and overheating protection for operation with non-flammable liquids, audible and visual alarm signals.



Reserve technical changes

### Options included

- Integrated Wi-Fi communication: The device may only be imported and used in certified countries and regions: United States of America, Canada, EU including the UK and Switzerland, and India.

LAUDA DR. R. WOBSEY GMBH & CO. KG  
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0  
info@lauda.de • www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim • HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser

## PRODUCT DATA SHEET

Published: 2026-04-28

LAUDA Universa U 830 E  
Cooling thermostat 100 V; 50/60 Hz  
Part Number: L003978



Working temperature min.  
-30 °C



Working temperature max.  
100 °C

### Technical Features

Working temperature range	-30 ... 100 °C
Operating temperature range	-30 ... 100 °C
Ambient temperature range	5 ... 40 °C
Temperature stability	0.05 ± K
Heating capacity range	1.0 ... 1.0 kW
Power consumption max.	1.1 kW
Current max.	11 A
Size of bath (W x D x H)	150 x 150 x 200 mm
Bath opening (WxD)	150 x 150 mm
Bath volume min. / max.	6.4 / 8.5 L
Pump Pressure max.	0,2 bar
Pump flow rate max. (pressure)	15 L/min
Overall dimensions (WxDxH)	260 x 480 x 630 mm
Weight	28 kg
Refrigerant stage 1	R-600a (GWP 3); 0.030 kg; 0.0 t CO <sub>2</sub> -eq
Power supply	100 V; 50/60 Hz
Power plug	Power cord with plug (NEMA 5-15P)

Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG  
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0  
info@lauda.de • www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim • HRA 560069


Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim • HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser

## PRODUCT DATA SHEET

Published: 2026-04-28

LAUDA Universa U 830 E  
 Cooling thermostat 100 V; 50/60 Hz  
 Part Number: L003978



Power cord with plug (NEMA 5-15P)

Temperature	Heat transfer liquid	Cooling Capacity 50 Hz	Cooling Capacity 60 Hz
20 °C	Ethanol	0.3 kW	0.3 kW
10 °C	Ethanol	0.275 kW	0.275 kW
0 °C	Ethanol	0.23 kW	0.23 kW
-10 °C	Ethanol	0.175 kW	0.175 kW
-20 °C	Ethanol	0.1 kW	0.1 kW
-30 °C	Ethanol	0.01 kW	0.01 kW

Reserve technical changes

### Standard accessories

- 1 Bath cover

LAUDA DR. R. WOBSEY GMBH & CO. KG  
 Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0  
 info@lauda.de • www.lauda.de  
 WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
 Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
 LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
 Sitz Lauda-Königshofen  
 Registergericht Mannheim • HRB 560226

Geschäftsführer:  
 Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
 Dr. Marc Stricker  
 Beirat: Dr. Gerhard Wobser


## PRODUCT DATA SHEET

Published: 2026-04-28

LAUDA Universa U 830 E

Cooling thermostat 100 V; 50/60 Hz

Part Number: L003978



Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG  
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0  
info@lauda.de • www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim • HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser


## PRODUCT DATA SHEET

Published: 2026-04-28

LAUDA Universa U 830 E

Cooling thermostat 100 V; 50/60 Hz

Part Number: L003978



Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG  
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0  
info@lauda.de • www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim • HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser