

PRODUCT DATA SHEET

Standing: 2025-04-29

LAUDA Variocool VC 1200 NRTL

200-230 V; 50/60 Hz

Part Number: L002818

Features

- Process thermostat suitable for use with non-flammable heat transfer liquids
- Coloured TFT display for simultaneous indication of actual & set values and graphic illustration of the temperature profile
- Clear text menu navigation, six selectable languages DE, EN, FR, ES, IT, RU
- Easy input via cursor and soft keys
- SmartCool system for energy-saving digital cooling management including compressor on-off control
- Fully electronic continuous controller with PID action
- Electronic level indication and low level alarm
- USB and Ethernet interface equipped as standard
- Remote fault indication through floating contact
- Slots for interface modules on the backside of the device
- Option for upgrading up to 2 additional interfaces (RS 232/485, Profibus, analogue, contact or EtherCAT module)
- Upgradeable with Pt 100/LiBus module for external control and remote control via Command Touch console
- Integrated programmer with max. 150 segments, splittable in 5 programmes
- Integrated web server for browser based operation in local area networks via PC, tablet or smart phone, secure data transfer due to authentication procedure and encryption
- Powerful pressure pump
- Adjustable bypass for pressure limiting
- Digital display of pump pressure
- Filler opening on top, drain tap on the backside
- Management of heat transfer liquids with stored properties
- Operates with non flammable liquids (water, water/glycol)
- Condenser cooling Air
- IEC panel connector (C20) for connecting a mains cable



Reserve technical changes



Working temperature min.
-20 °C



Working temperature max.
80 °C

LAUDA DR. R. WOBSEY GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0
info@lauda.de • www.lauda.de
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser

PRODUCT DATA SHEET

Standing: 2025-04-29

LAUDA Variocool VC 1200 NRTL

200-230 V; 50/60 Hz

Part Number: L002818

Technical Features (according to DIN 12876)

Working temperature range	-20 ... 80 °C
Ambient temperature range	5 ... 40 °C
Temperature stability	0.05 ± K
Heater power min.	1.9 kW
Heater power max.	2.5 kW
Current max.	16 A
Pump Pressure max. 50 Hz	3.1 bar
Flow rate max. (pressure) 50 Hz	37 L/min
Pressure pump max. 60 Hz	4.1 bar
Flow rate max. (pressure) 60 Hz	44 L/min
In / Outlet connection thread (outside)	G 3/4"
Pressure adjustment	bypass
Filling volume min.	9 L
Filling volume max.	14 L
Refrigerant stage 1	R-449A (GWP 1397); 0.500 kg; 0.7 t CO ₂ -eq
International protection	IP 32
Noise level	59 dB(A)
Overall dimensions (WxDxH)	450 x 550 x 825 mm
Weight	65 kg
Power supply	200-230 V; 50/60 Hz
Power plug	Cold device plug (IEC 60320, C20)
Certification	NRTL according to UL 1995:2018, CSA C22.2 No.236:2015

Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0
info@lauda.de • www.lauda.de
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser

PRODUCT DATA SHEET

Standing: 2025-04-29

LAUDA Variocool VC 1200 NRTL

200-230 V; 50/60 Hz

Part Number: L002818

Temperature	Heat transfer liquid	Cooling Capacity 50 Hz	Cooling Capacity 60 Hz
80 °C	Waterglycol	1.4 kW	1.55 kW
20 °C	Ethanol	1.55 kW	1.65 kW
10 °C	Ethanol	1.15 kW	1.25 kW
0 °C	Ethanol	0.75 kW	0.8 kW
-10 °C	Ethanol	0.45 kW	0.45 kW
-20 °C	Ethanol	0.18 kW	0.14 kW

Standard accessories

- 2 nipples 3/4" with screw cap G3/4 for pump connectors
- 1 Mains cable (C19) without plug, length: 6 meters
- 1 Mains plug, Schuko (CEE7/7)
- 1 Mains plug, NEMA 6-20P

Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0
info@lauda.de • www.lauda.de
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser


PRODUCT DATA SHEET

Standing: 2025-04-29

LAUDA Variocool VC 1200 NRTL

200-230 V; 50/60 Hz

Part Number: L002818



Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG
 Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0
 info@lauda.de • www.lauda.de
 WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
 Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
 LAUDA DR. R. WOBSEY Verwaltungs-GmbH
 Sitz Lauda-Königshofen
 Registergericht Mannheim • HRB 560226

Geschäftsführer:
 Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
 Dr. Marc Stricker
 Beirat: Dr. Gerhard Wobser