

PRODUCT DATA SHEET

Published: 2026-03-19

LAUDA Variocool VC 1200 W

Process thermostat 230 V; 50 Hz

Part Number: L004617

Features

- Coloured TFT display for simultaneous indication of actual & set values and graphic illustration of the temperature profile
- Clear text menu navigation, six selectable languages DE, EN, FR, ES, IT, RU
- Easy input via cursor and soft keys
- Fully electronic continuous controller with PID action
- Electronic level indication and low level alarm
- Powerful pressure pump
- USB interface as standard
- Remote fault indication through floating contact
- Upgradeable with an interface module (analogue module, contact module, RS 232/485 module, Profibus, Ethernet-USB module)
- Integrated programmer with max. 150 segments, splittable in 5 programmes
- Adjustable bypass for pressure limiting
- Filler opening on top, drain tap on the backside
- SmartCool system for energy-saving digital cooling management including compressor on-off control
- Operates with non flammable liquids (water, water/glycol)
- Process thermostat suitable for use with non-flammable heat transfer liquids
- Condenser cooling Water
- Utilizes natural refrigerants



Reserve technical changes



Working temperature min.
-25 °C



Working temperature max.
80 °C

LAUDA DR. R. WOBSEY GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0
info@lauda.de • www.lauda.de
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser

PRODUCT DATA SHEET

Published: 2026-03-19

LAUDA Variocool VC 1200 W

Process thermostat 230 V; 50 Hz

Part Number: L004617

Technical Features (according to DIN 12876)

| | |
|---|--|
| Working temperature range | -25 ... 80 °C |
| Ambient temperature range | 5 ... 40 °C |
| Temperature stability | 0.05 ± K |
| Heater power max. | 2.3 kW |
| Power consumption max. | 2.7 kW |
| Current max. | 12 A |
| Pump Pressure max. | 0,9 bar |
| Pump flow rate max. (pressure) | 28 L/min |
| In / Outlet connection thread (outside) | G 3/4" |
| Pressure adjustment | bypass |
| Filling volume min. | 8 L |
| Filling volume max. | 15 L |
| Recommended cooling water temperature | 15 °C |
| Cooling water consumption | 3 L/min |
| Pressure difference cooling water min. | 1 bar |
| Maximum pressure cooling water | 10 bar |
| Overall dimensions (WxDxH) | 450 x 550 x 650 mm |
| Weight | 50 kg |
| Refrigerant stage 1 | R-290 (GWP 3); 0.095 kg; 0.0 t CO ₂ -eq |
| Power supply | 230 V; 50 Hz |
| Power plug | Power cord with plug (SEV 1011, SEV 5934/2, T23) |

Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0
info@lauda.de • www.lauda.de
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser


PRODUCT DATA SHEET

Published: 2026-03-19

LAUDA Variocool VC 1200 W

Process thermostat 230 V; 50 Hz

Part Number: L004617



Power cord with plug (SEV 1011, SEV 5934/2, T23)

| Temperature | Heat transfer liquid | Cooling Capacity 50 Hz |
|-------------|----------------------|------------------------|
| 20 °C | Ethanol | 1.4 kW |
| 10 °C | Ethanol | 1.3 kW |
| 0 °C | Ethanol | 1.1 kW |
| -10 °C | Ethanol | 0.85 kW |
| -20 °C | Ethanol | 0.53 kW |
| -25 °C | Ethanol | 0.34 kW |

Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0
info@lauda.de • www.lauda.de
WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser


PRODUCT DATA SHEET

Published: 2026-03-19

LAUDA Variocool VC 1200 W

Process thermostat 230 V; 50 Hz

Part Number: L004617



Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0
info@lauda.de • www.lauda.de
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser