

PRODUCT DATA SHEET

Integral IN 530 TW

The new generation of Integral T process thermostats for professional external temperature control from -30 to 120 °C: The new LAUDA Integral T process thermostats ensure efficient control of external temperature control processes in the temperature range from -30 to 150 °C. The units enable rapid temperature changes due to adapted heating and cooling capacities with small internal volumes. Due to the open hydraulic system, the unit vents quickly and without functional restrictions. This makes it ideal for temperature control processes with frequent consumer or specimen changes. Classic areas of application are reaction controls or climate simulations. A reliable, powerful submersible pump and the internal bypass for pressure limitation are standard features of the Integral T.



Working temperature min.
-30 °C



Working temperature max.
120 °C

Technical Attributes

Integral IN 530 TW

Working temperature min.	-30 °C
Working temperature max.	120 °C
Ambient temperature min.	5 °C
Ambient temperature max.	40 °C
Temperature stability	0.05 ±K
Filling volume min.	7.2 L
Filling volume max.	20.5 L
Heater power	8.0 kW
Cooling output at 20°C measured with ethanol	6.0 kW
Cooling output at 10°C measured with ethanol	5.5 kW
Cooling output at 0°C measured with ethanol	4.5 kW
Cooling output at -10°C measured with ethanol	3.0 kW
Cooling output at -20°C measured with ethanol	1.6 kW
Cooling output at -30°C measured with ethanol	0.7 kW
Power consumption	11.0 kW

LAUDA DR. R. WOBSEY GMBH & CO. KG
Pfarrstraße 41/43 · 97922 Lauda-Königshofen
Postfach 1251 · 97912 Lauda-Königshofen · DE

T +49 (0) 9343 503-0 · F +49 (0) 9343 503-222
info@lauda.de · www.lauda.de
WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim · HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim · HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser

PRODUCT DATA SHEET

Integral IN 530 TW

Technical Attributes	Integral IN 530 TW
Tube size	$\frac{3}{4}$ "
Pump connection thread	G $\frac{3}{4}$
Pump pressure max.	3.5 bar
Pump flow max. (pressure)	40 L/min
Dimensions (WxDxH) in mm	560x550x1325

LAUDA DR. R. WOBSEY GMBH & CO. KG
 Pfarrstraße 41/43 · 97922 Lauda-Königshofen
 Postfach 1251 · 97912 Lauda-Königshofen · DE

T +49 (0) 9343 503-0 · F +49 (0) 9343 503-222
 info@lauda.de · www.lauda.de
 WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
 Registergericht Mannheim · HRA 560069

Persönlich haftende Gesellschafterin:
 LAUDA DR. R. WOBSEY Verwaltungs-GmbH
 Sitz Lauda-Königshofen
 Registergericht Mannheim · HRB 560226

Geschäftsführer:
 Dr. Gunther Wobser (Vors.), Dr. Mario Englert
 Dr. Marc Stricker
 Beirat: Dr. Gerhard Wobser