

PRODUCT DATA SHEET

Integral IN 950 XTW

The new generation of the successful Integral XT process thermostats for professional temperature control in the temperature range from -90 to 320 °C: The powerful Integral XT process thermostats use the flow principle with cold oil superimposition. It allows users to use a wide temperature range with one temperature control medium. The electronically controlled, magnetically coupled eight-stage LAUDA Variopump allows the optimum thermal connection of the flow rate both for pressure-sensitive consumers and for applications with high hydraulic resistance. The modular interface concept ensures maximum networking of the user processes.



Working temperature min.
-50 °C



Working temperature max.
220 °C

Technical Attributes

| Technical Attributes | Integral IN 950 XTW |
|---------------------------------------------------|---------------------|
| Working temperature min. | -50 °C |
| Working temperature max. | 220 °C |
| Ambient temperature min. | 5 °C |
| Ambient temperature max. | 40 °C |
| Temperature stability | 0.05 ±K |
| Application | external |
| Filling volume min. | 4.8 L |
| Filling volume max. | 17.2 L |
| Heater power | 8.0 kW |
| Cooling output at 200°C measured with thermal oil | 9.5 kW |
| Cooling output at 100°C measured with thermal oil | 9.5 kW |
| Cooling output at 20°C measured with ethanol | 9.5 kW |
| Cooling output at 10°C measured with ethanol | 8.5 kW |
| Cooling output at 0°C measured with ethanol | 6.2 kW |
| Cooling output at -10°C measured with ethanol | 4.3 kW |

LAUDA DR. R. WOBSEY GMBH & CO. KG
Pfarrstraße 41/43 · 97922 Lauda-Königshofen
Postfach 1251 · 97912 Lauda-Königshofen · DE

T +49 (0) 9343 503-0 · F +49 (0) 9343 503-222
info@lauda.de · www.lauda.de
WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim · HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim · HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser

PRODUCT DATA SHEET

Integral IN 950 XTW

| Technical Attributes | Integral IN 950 XTW |
|-----------------------------------------------|---------------------|
| Cooling output at -20°C measured with ethanol | 3.0 kW |
| Cooling output at -30°C measured with ethanol | 1.7 kW |
| Cooling output at -40°C measured with ethanol | 0.9 kW |
| Cooling output at -50°C measured with ethanol | 0.35 kW |
| Power consumption | 12.0 kW |
| Interface(s) | Ethernet, USB |
| Noise level | 66 dB(A) |
| Pump connection thread | M30 x 1,5 |
| Pump pressure max. | 3.1 bar |
| Pump flow max. (pressure) | 65 L/min |
| Dimensions (WxDxH) in mm | 560x550x1325 |

LAUDA DR. R. WOBSEY GMBH & CO. KG
Pfarrstraße 41/43 · 97922 Lauda-Königshofen
Postfach 1251 · 97912 Lauda-Königshofen · DE

T +49 (0) 9343 503-0 · F +49 (0) 9343 503-222
info@lauda.de · www.lauda.de
WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim · HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim · HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser