

## PRODUCT DATA SHEET

### Integral IN 280 XTW

The new generation of the successful Integral XT process thermostats for professional temperature control in the temperature range from -90 to 320 °C: The powerful Integral XT process thermostats use the flow principle with cold oil superimposition. It allows users to use a wide temperature range with one temperature control medium. The electronically controlled, magnetically coupled eight-stage LAUDA Variopump allows the optimum thermal connection of the flow rate both for pressure-sensitive consumers and for applications with high hydraulic resistance. The modular interface concept ensures maximum networking of the user processes.



Working temperature min.  
-80 °C



Working temperature max.  
220 °C

#### Technical Attributes

Working temperature min.	-80 °C
Working temperature max.	220 °C
Ambient temperature min.	5 °C
Ambient temperature max.	40 °C
Temperature stability	0.05 ±K
Application	external
Filling volume min.	4.8 L
Filling volume max.	17.2 L
Heater power	4.0 kW
Cooling output at 200°C measured with thermal oil	1.70 kW
Cooling output at 100°C measured with thermal oil	1.70 kW
Cooling output at 20°C measured with ethanol	1.70 kW
Cooling output at 10°C measured with ethanol	1.65 kW
Cooling output at 0°C measured with ethanol	1.60 kW
Cooling output at -10°C measured with ethanol	1.60 kW

#### Integral IN 280 XTW

-80 °C
220 °C
5 °C
40 °C
0.05 ±K
external
4.8 L
17.2 L
4.0 kW
1.70 kW
1.70 kW
1.70 kW
1.65 kW
1.60 kW
1.60 kW

LAUDA DR. R. WOBSER GMBH & CO. KG  
Pfarrstraße 41/43 · 97922 Lauda-Königshofen  
Postfach 1251 · 97912 Lauda-Königshofen · DE

T +49 (0) 9343 503-0 · F +49 (0) 9343 503-222  
info@lauda.de · www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim · HRB 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSER Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim · HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser

## PRODUCT DATA SHEET

### Integral IN 280 XTW

#### Technical Attributes

Cooling output at -20°C measured with ethanol
Cooling output at -30°C measured with ethanol
Cooling output at -40°C measured with ethanol
Cooling output at -50°C measured with ethanol
Cooling output at -60°C measured with ethanol
Cooling output at -70°C measured with ethanol
Cooling output at -80°C measured with ethanol
Power consumption
Interface(s)
Noise level
Pump connection thread
Pump pressure max.
Pump flow max. (pressure)
Dimensions (WxDxH) in mm

#### Integral IN 280 XTW

1.80 kW
1.80 kW
1.80 kW
1.50 kW
0.90 kW
0.45 kW
0.18 kW
9.0 kW
Ethernet, USB
64 dB(A)
M30 x 1,5
3.1 bar
65 L/min
560x550x1325

LAUDA DR. R. WOBSER GMBH & CO. KG  
 Pfarrstraße 41/43 · 97922 Lauda-Königshofen  
 Postfach 1251 · 97912 Lauda-Königshofen · DE

T +49 (0) 9343 503-0 · F +49 (0) 9343 503-222  
 info@lauda.de · www.lauda.de  
 WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
 Registergericht Mannheim · HRA 560069

Persönlich haftende Gesellschafterin:  
 LAUDA DR. R. WOBSER Verwaltungs-GmbH  
 Sitz Lauda-Königshofen  
 Registergericht Mannheim · HRB 560226

Geschäftsführer:  
 Dr. Gunther Wobser (Vors.), Dr. Mario Englert  
 Dr. Marc Stricker  
 Beirat: Dr. Gerhard Wobser