

PRODUCT DATA SHEET



Integral XT 1850 WS

The Integral XT process thermostats operate according to the flow principle with a cold oil blanket that allows the utilization of the entire temperature range through the use of a heat transfer medium. The electronically controlled, magnetically coupled pump can alter the pressure to adapt the flow rate to the relevant process requirements. The Integral XT models can be integrated easily in different process control systems using a wide selection of interface modules.



Working temperature min.
-50 °C



Working temperature max.
220 °C

Technical Attributes	Integral XT 1850 WS
Working temperature min.	-50 °C
Working temperature max.	220 °C
Ambient temperature min.	5 °C
Ambient temperature max.	40 °C
Temperature stability	0.30 ±K
Application	external
Filling volume min.	9.00 L
Filling volume expansion vessel	17.4 L
Heater power	16.00 kW
Cooling output at 200°C measured with thermal oil	18.50 kW
Cooling output at 150°C measured with thermal oil	18.50 kW
Cooling output at 100°C measured with thermal oil	18.50 kW
Cooling output at 50°C measured with thermal oil	18.30 kW
Cooling output at 20°C measured with ethanol	18.50 kW
Cooling output at 10°C measured with ethanol	12.50 kW

LAUDA DR. R. WOBSEY GMBH & CO. KG
Pfarrstraße 41/43 · 97922 Lauda-Königshofen
Postfach 1251 · 97912 Lauda-Königshofen · DE

T +49 (0) 9343 503-0 · F +49 (0) 9343 503-222
info@lauda.de · www.lauda.de
WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim · HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim · HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser

PRODUCT DATA SHEET

Integral XT 1850 WS

Technical Attributes	Integral XT 1850 WS
Cooling output at 0°C measured with ethanol	10.30 kW
Cooling output at -10°C measured with ethanol	7.70 kW
Cooling output at -20°C measured with ethanol	5.90 kW
Cooling output at -30°C measured with ethanol	3.80 kW
Cooling output at -40°C measured with ethanol	2.20 kW
Cooling output at -50°C measured with ethanol	1.20 kW
Power consumption	17.30 kW
Interface(s)	RS-232/485
Pump connection thread	M38 x 1,5 A (DN 25)
Pump pressure max.	5.8 bar
Pump flow max. (pressure)	90 L/min
Dimensions (WxDxH) in mm	700x550x1600

LAUDA DR. R. WOBSEY GMBH & CO. KG
 Pfarrstraße 41/43 · 97922 Lauda-Königshofen
 Postfach 1251 · 97912 Lauda-Königshofen · DE

T +49 (0) 9343 503-0 · F +49 (0) 9343 503-222
 info@lauda.de · www.lauda.de
 WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
 Registergericht Mannheim · HRA 560069

Persönlich haftende Gesellschafterin:
 LAUDA DR. R. WOBSEY Verwaltungs-GmbH
 Sitz Lauda-Königshofen
 Registergericht Mannheim · HRB 560226

Geschäftsführer:
 Dr. Gunther Wobser (Vors.), Dr. Mario Englert
 Dr. Marc Stricker
 Beirat: Dr. Gerhard Wobser